
Multi-dimensional measuring instruments

Part 4: Type evaluation report format

Instruments de mesure multidimensionnels

Partie 4: Format du rapport d'évaluation de type



Contents

Foreword.....	4
Introduction	5
Explanatory notes	6
Identification of the instrument	7
General information concerning the type	10
Configuration for type evaluation.....	13
Selection of sample(s)	14
Adjustments or modifications.....	15
Summary of test report(s).....	16
Appraisal of test data (where applicable)	17
Summary of the checklist	18
Checklist.....	20

Foreword

The International Organisation of Legal Metrology (OIML) is a worldwide, intergovernmental organisation whose primary aim is to harmonise the regulations and metrological controls applied by the national metrological services, or related organisations, of its Member States. The main categories of OIML publications are:

- **International Recommendations (OIML R)**, which are model regulations that establish the metrological characteristics required of certain measuring instruments and which specify methods and equipment for checking their conformity. OIML Member States shall implement these Recommendations to the greatest possible extent;
- **International Documents (OIML D)**, which are informative in nature and which are intended to harmonise and improve work in the field of legal metrology;
- **International Guides (OIML G)**, which are also informative in nature and which are intended to give guidelines for the application of certain requirements to legal metrology; and
- **International Basic Publications (OIML B)**, which define the operating rules of the various OIML structures and systems.

OIML Draft Recommendations, Documents and Guides are developed by Project Groups linked to Technical Committees or Subcommittees which comprise representatives from the Member States. Certain international and regional institutions also participate on a consultation basis. Cooperative agreements have been established between the OIML and certain institutions, such as ISO and the IEC, with the objective of avoiding contradictory requirements. Consequently, manufacturers and users of measuring instruments, test laboratories, etc. may simultaneously apply OIML publications and those of other institutions.

International Recommendations, Documents, Guides and Basic Publications are published in English (E) and translated into French (F) and are subject to periodic revision.

Additionally, the OIML publishes or participates in the publication of **Vocabularies (OIML V)** and periodically commissions legal metrology experts to write **Expert Reports (OIML E)**. Expert Reports are intended to provide information and advice, and are written solely from the viewpoint of their author, without the involvement of a Technical Committee or Subcommittee, nor that of the CIML. Thus, they do not necessarily represent the views of the OIML.

This publication – reference OIML R 129-4, Edition 2020 – was developed by Project Group 1 of OIML TC 7/SC 5 *Dimensional measuring instruments*. It was approved for final publication by the International Committee of Legal Metrology in 2020. It supersedes the previous edition dated 2000.

OIML Publications may be downloaded from the OIML web site in the form of PDF files. Additional information on OIML Publications may be obtained from the Organisation's headquarters:

Bureau International de Métrologie Légale
11, rue Turgot - 75009 Paris - France
Telephone: 33 (0)1 48 78 12 82
Fax: 33 (0)1 42 82 17 27
E-mail: biml@oiml.org
Internet: www.oiml.org

Introduction

This “Type evaluation report format”, the subject of OIML R 129-4, aims at presenting, in a standardised format, the results of the evaluation to which a type of a multi-dimensional measuring instrument shall be submitted with a view to its approval.

The “Checklist” is a summary of the evaluation and examinations carried out on the instrument. It includes the conclusions of the results of the tests performed, experimental or visual checks based on the required performance criteria and associated tests in OIML R 129-1 and -2. The words or condensed sentences intend to remind the examiner of the requirements of R 129-1 and -2 without reproducing them.

All metrology services evaluating types of multi-dimensional measuring instrument according to OIML R 129-1 and -2 or to national or regional regulations based on OIML R 129-1 and -2 are strongly advised to use this “Type evaluation report format”, directly or after translation into a language other than English or French. Its direct use in English or in French, or in both languages, is even more strongly recommended whenever the results of type evaluation may be transmitted by the country performing these evaluations to the approving authorities of another country, under bi- or multi-lateral cooperation agreements. In the framework of the OIML Certification System (OIML-CS), use of the “Type evaluation report format” is mandatory.

Explanatory notes

Symbols	Meaning
L	Indicated length
W	Indicated width
H	Indicated height
L_T	Length of the test object
ΔL	Error, $L - L_T$
W_T	Width of the test object
ΔW	Error, $W - W_T$
H_T	Height of the test object
ΔH	Error, $H - H_T$
mpe	Maximum permissible error
V	Volume indicated on the instrument
V_{calc}	$L \times W \times H$
F	Conversion factor
DW	Dimensional weight indicated on the instrument
DW_{calc}	$V \times F$
SF	Significant fault

Identification of the instrument

Application no.:

Type designation:

Identification no.:

Manufacturer:

Software version:

Report date:

Documentation from the manufacturer

(Record as necessary to identify the equipment under evaluation)

System or module name	Drawing number or software reference	Issue level	Document identification
.....			
.....			
.....			
.....			
.....			
.....			
.....			

Simulator documentation (if applicable)

System or module name	Drawing number or software reference	Issue level	Document identification
.....			
.....			
.....			
.....			
.....			
.....			
.....			

Identification of the instrument (continued)

Application no.:		Type designation:	
Identification no.:		Manufacturer:	
Report date:			

Simulator function (summary) (if applicable)

(Simulator description and drawings, block diagram, etc. should be attached to the report if available)

Identification of the instrument (continued)

Application no.:		Type designation:	
Identification no.:		Manufacturer:	
Report date:			

Physical description of the instrument

Describe, using point form, the physical construction of the MDMI (materials, configuration and location of components, interfaces and communications ports). Attach photographs, diagrams or drawings if available:

Describe, using point form, the measurement technology used (include details such as physical contact method, laser class, power and wavelength, ultrasonic frequency, or camera spectrum):

General information concerning the type

Application no.: Manufacturer:

Type designation: Applicant:

Instrument category:

Parameter limits		
	Minimum	Maximum
Temperature limits (°C)		
Speed limitations (m/s)		
Voltage (V)		
Minimum spacing		/

Limitation	Check if applicable
Object	
Cuboidal	
Irregular	
Singulated	
Non-singulated, non-touching	
Touching	
Measurement dynamics	
Static measurement only	
Fixed speed	
Variable speed	
Unidirectional	
Bidirectional	
Installation	
Permanent	
Mobile	
Power supply	
AC	
AC-DC converter	
Battery	
DC	

Evaluation period:

Date of report:

Observer:

General information concerning the type (continued)

Application no.: Manufacturer:
 Type designation: Applicant:
 Instrument category:

Scale interval and limits of indication

Axis	Unit of measurement	Scale interval (d)	Minimum dimension	Maximum dimension
X				
Y				
Z				

Describe, using point form, each axis and its relation to the physical object being measured and/or the MDMI itself:

Use this space to describe, using point form, any other use or installation limitations not detailed in the above on the previous page (such as special applications other than postage, freight or storage, restrictions on object material, texture, reflectivity or colour, object positioning):

Use this space to indicate additional remarks and/or information: connecting equipment, interfaces, choice of the manufacturer regarding protection against disturbances, etc.:

General information concerning the type (continued)

Application no.:		Manufacturer:	
Type designation:		Applicant:	
Instrument category:			

Indications and controls

Describe, using point form, all indications and controls of the instrument (such as wired or wireless communication with the instrument, zero method, ready indication, computed quantities, error codes). Describe each measurement (*L*, *W*, and *H*) and its relation to the physical object being measured and/or the MDMI itself:

Sealing

Describe, using point form, the physical and electronic seals (e.g. audit trails) used to protect the metrological characteristics of the instrument, and how to access them. Also describe any remote access abilities available and how these are sealed:

Software

Describe, using point form, the means used to protect legally relevant software in the instrument and indicate the version of the software present at the time of testing and how to verify this version number:

Configuration for type evaluation

Application no.:		Type designation:	
Report date:		Manufacturer:	

Use this space for additional information relating to equipment configuration, interfaces, data rates, EMC protection options, etc., for the instrument and/or simulator:

Selection of sample(s)

Application no.:

Type designation:

Report date:

Manufacturer:

Use this space for additional information relating to the justification for the selection of sample(s), in particular if a family of instruments or modules or if specific requirements are mentioned in OIML R 129-1 and -2:

Adjustments or modifications

Application no.: Type designation:
Report date: Manufacturer:

Use this space for additional information relating to the identification of any authorised and agreed upon adjustments or modifications made to the sample or samples during the evaluation:

Summary of test report(s)

Use the table below to summarise the test report(s) used to support the type evaluation:

Test report number	Issued by	Remarks*

* Use this column to record if the test report was issued:

- under the OIML Basic Certificate System, the OIML Mutual Acceptance Arrangement (MAA) or the OIML Certification System (OIML-CS) Scheme A or B. Where the test report was used as the basis for issuing an existing OIML Certificate, the relevant OIML Certificate number should be noted;
- by a Manufacturer Test Laboratory (MTL);
- under the scope of ISO/IEC 17025 accreditation.

Appraisal of test data (where applicable)

Use this space to record the appraisal of test data:

Requirement	Check if yes	Remarks (provide information)
Correct method has been used		
Test data from acceptable testing facility		
Testing facility capable of performing the testing (accredited or peer assessed)		
Test report provided in OIML format		
Test performed against OIML R 129 (versions)		
Results acceptable for all the tests performed		
Further information/clarification needed		
Further testing to be performed		

If there is more than one test report, indicate in the remarks if the answers are not the same for all test reports by referring to the test report number from the previous table.

Corrective actions required

Issue identified	Details of corrective action required

Summary of the checklist

For each test, the “Summary of the checklist” below and the “Checklist” in clause 1 shall be completed according to this example:

	Passed	Failed
When the instrument has passed the test:	X	
When the instrument has failed the test:		X
When the test is not applicable:	/	/

Summary of the checklist:

Requirement	Passed	Failed	Remarks
Metrological requirements R 129-1 clause 4			
Technical requirements R 129-1 clause 5			
Additional requirements for software controlled multi-dimensional measuring instruments R 129-1 clause 6			
Test procedures R 129-2			
Overall result			

Application no.:		Type designation:	
Report date:		Manufacturer:	

Use this page to detail remarks from the summary of the checklist:

Checklist

Report no:
 Application no:
 Manufacturer:
 Make and model:
 Serial no:
 Date:
 Observer:

Requirement (R129-1)		Passed	Failed	Remarks
3	Units of measurement			
	Correct units and symbols used			
4.1.1	Scale intervals, minimum dimension			
	Correct minimum dimensions			
4.1.2	Value of mpe			
	mpe applicable to any of the three dimensions = $\pm 1.0 d$			
4.1.3	Value of the fault limit			
	Value of one scale interval (d)			
4.1.4	Variation between indicators			
	No variation in indications			
4.1.5	Multi-interval instruments			
	Appropriate mpe for applicable range and axis			
4.1.6	Calculated quantities			
	Derived from displayed value of measured dimensions			
	Rounded to nearest applicable scale interval			
	Calculated quantities in mathematical agreement			
4.2.1	Rated operating conditions			
	At least 30 °C			
	If powered by direct current, instrument functions correctly or does not indicate quantity below manufacturer's specified nominal voltage			
	Instrument designed and manufactured such that it does not exceed the mpe when exposed to:			
4.2.1 (a)	– mains power voltage variations: – 15 % to + 10 % of nominal voltage			
4.2.1 (b)	– air temperature variations at the temperature limits stated in the descriptive markings or – 10 °C to + 40 °C			
4.2.1 (c)	– relative humidity of 85 % (non-condensing) at the high temperature limit of the instrument or 40 °C, whichever is lower			
4.3.1	Disturbances applied to measuring instrument			
	Instrument is designed and manufactured such that, when exposed to disturbances, either:			
4.3.1 (a)	– significant fault do not occur, or			
4.3.1 (b)	– significant faults are detected and acted upon			
4.3.3	Tests for disturbances; severity levels			
	Instrument withstands disturbances when subject to the applicable severity levels			
4.4	Light and acoustic effects			
	Remains within mpe when subjected to the applicable light or acoustic effects, or			
	has provisions for alternative operations			
5.1.1	Fraudulent use			
	Instrument shall not facilitate fraudulent use			

Checklist (continued)

Requirement (R129-1)		Passed	Failed	Remarks
5.1.2	Suitability of construction			
	All controls, indicators etc. are suitable			
5.1.3	Suitability for use			
	Suit the method of operation and objects			
5.1.4	Suitability for verification			
	Constructed so that test of performance requirements can be carried out			
	Test mode provided (only volume indicated in normal position)			
5.1.5	Zero or ready adjustment			
	Facilitates for zero or ready condition			
	Can only be set with no object in the measurement area			
	Zero or ready condition indicated			
	Condition set automatically or inhibited if not set correctly			
5.1.6	Tare device			
5.1.6 (a)	Only operates subtractively with respect to the zero or ready condition			
5.1.6 (b)	Value of the tare scale interval is the same as that for the respective axis and range			
5.1.6 (c)	Operation of tare indicated			
5.1.7	Warm-up			
	Indications within mpe following warm-up			
5.2.1	Indicators and printing devices			
5.2.1 (a)	Instrument has an indicator or printer which displays dimensions or volume			
5.2.1 (b)	Device to transmit, store and preserve measurement results			
5.2.1 (c)	Indications automatically displayed or are readily available			
5.2.1 (d)	Other indications (e.g. DW,F) are automatically displayed or are readily available			
5.2.1 (e)	Previously displayed indication persists long enough for easy reading by observer			
5.2.1 (f)	Indication clearly assignable to a specific object			
5.2.1 (g)	Display of extended indication device:			
	– while pressing a key; or			
	– limited to 5 seconds			
	Printing and data transmission restricted in extended indication			
5.2.1 (h)	Extended indication device not fitted to instrument for direct sales to public			
5.2.1 (i)	All indications are identified			
5.2.2	Presentation of indications			
	Indications, printing reliable, clear and unambiguous and printing indelible			
	Figures easy to read			
	Printed numbers and symbols at least 2 mm high			
	Digits oriented normally and permit reading by simple juxtaposition			
5.2.3	Units of measurement			
	All indications include the name/symbol of the unit of measurement			
	On tickets, name or symbol printed by printer or pre-printed			
	For any one indication, only one unit of measurement used			
5.2.4	Value of scale interval			
	Value of Scale interval in the form $1, 2 \text{ or } 5 \times 10^n$			
	The scale interval shall be:			
5.2.4 (a)	– the same for each axis; or			
5.2.4 (b)	– different for one axis from the other provided instructions are marked, or indication of incorrect use given; or			

Checklist (continued)

Requirement (R129-1)		Passed	Failed	Remarks
5.2.4 (c)	– variable, on one or more axes provided:			
	– all three axes are multi-interval - all the same			
	– two axes are multi-interval and the third is fixed			
	– instrument limitations are clearly marked			
	– one axis is multi-interval and the others are fixed			
	– instrument limitations are clearly marked			
5.2.5	Decimal numbers			
	At least one zero before decimal mark for values < 1			
	Decimal mark printed			
	One or more fixed zeros to the right of variable numbers for values > 1			
	All decades to right of decimal are active and least significant digit is scale interval			
5.2.6	Limits of indication			
	Displaying, storing, transmitting or printing the quantity value is inhibited if the entire object is not in, or does not pass through, the measuring area			
5.2.6 (a)	Dimensions below minimum either:			
	– blank; or			
	– displayed with error code			
5.2.6 (b)	Dimensions above maximum + 9 d either:			
	– blank; or			
	– displayed with error code			
5.2.7	Multi-interval instruments			
5.2.7	For each partial measuring range:			
5.2.7 (a)	– $d_1 < d_2 < \dots < d_r$;			
5.2.7 (b)	– Min = Min ₁ , Max = Max _r , Max ₁ = Min ₂ etc.			
5.2.7 (c)	– Minimum dimension of any axis = minimum dimension of the lowest partial range of that axis			
5.2.7 (d)	– Maximum dimension of any axis = minimum dimension of the highest partial range of that axis			
5.2.7 (e)	– Minimum dimension of any partial measuring range is no smaller than minimum specified in 4.1.1			
5.2.8	Multi-instrument systems			
5.2.8	Indications on common indicator and test indicator agree			
	Indication from each device clearly identified on the common indicator			
5.2.9	Printed and displayed information - general			
5.2.9.1	Displayed results include the measured dimensions			
5.2.9.2	Printed ticket or stored measurement includes:			
5.2.9.2 (a)	– dimensions: length (<i>L</i>), width (<i>W</i>) and height (<i>H</i>)			
5.2.9.2 (b)	– date, transaction number, etc.			
5.2.9.2 (c)	– volume (<i>V</i>)			
5.2.9.2 (d)	– weight (<i>Wt</i>)			
5.2.9.2 (e)	– dimensional weight (DW.....kg)			
5.2.9.2 (f)	– dimensional tare (DT.....kg) or linear tare (LT.....cm)			
5.2.9.2 (g)	– conversion factor (<i>F</i>)			
5.2.9.2 (h)	– quantity for charging			
5.2.9.2 (i)	– price rate and price			
5.2.9.3	Additional requirements for printed tickets			
	A printed ticket contains printed or pre-printed notices stating:			
5.2.9.3 (a)	– dimensions and/or volume are those of the smallest rectangular box			
5.2.9.3 (b)	– dimensional weight is a calculated value			

Checklist (continued)

Requirement (R129-1)		Passed	Failed	Remarks
5.2.10	Availability of indications			
	Indications must be available to the customer as follows:			
5.2.10 (a)	– if customer present, all indications are available to customer at time of measurement			
5.2.10 (b)	– if customer not present, information in 5.2.9 shall be available on request, if not displayed or printed			
5.3.1	Nameplate			
	Instrument or descriptive nameplate permanently affixed to the instrument, permanently and clearly marked and clearly visible at all times			
	Nameplate contains the following information			
5.3.1 (a)	– manufacturer's name or mark			
5.3.1 (b)	– model designation			
5.3.1 (c)	– serial number and year of manufacture			
5.3.1 (d)	– type evaluation mark			
5.3.1 (e)	– minimum and maximum dimensions for each axis			
5.3.1 (f)	– maximum and minimum measuring speeds			
5.3.1 (g)	– scale interval(s) in the form of $d =$			
5.3.1 (h)	– temperature limits (if other than -10 °C to 40 °C)			
5.3.2	Technical specifications			
	Specifications or limitations of use clearly marked and visible to the operator, or in operator's manual			
5.3.2 (a)	Special application			
5.3.2 (b)	Minimum spacing			
5.3.2 (c)	Measure only rectangular boxes			
5.3.2 (d)	Box location			
5.3.2 (e)	Limitations of surface characteristics			
	Other special notices relating to the instrument			
5.4.1	Verification mark			
	Provision made for the application of a verification mark			
	The following requirements apply:			
5.4.1 (a)	– mark easily affixed without affecting the metrological properties			
5.4.1 (b)	– mark visible without moving or dismantling instrument when in use			
5.4.1 (c)	the part on which the mark is located is not removable from the instrument without damaging the mark			
5.4.1 (d)	the size of the space sufficient for a mark (e.g. at least 200 mm^2)			
5.4.2	Sealing			
	Provision made for sealing by mechanical or electronic means			
	Mechanical seal applied			
	For electronic seals:			
5.4.2 (a)	– access by authorised persons protected by physical key or password			
5.4.2 (b)	– access to alter parameters automatically recorded			
5.4.2 (c)	– record readily accessible by simple action			
5.4.2 (d)	– record readily identifiable			
5.4.2 (e)	– reference record permanently marked on the instrument			
5.4.2 (f)	– record does not repeat in a sequence of less than 999 alterations			
	– record persists reliably for a period of at least two years			
5.4.2 (g)	– record persists through tests for influence factors and disturbances			

Checklist (continued)

Requirement (R129-1)		Passed	Failed	Remarks
5.5.2	Interfaces for peripheral devices			
	Interface does not allow metrological functions to be affected by the operation of the ancillary devices or connected instruments or disturbances acting on the interface			
	Interface sealed if instructions or data affecting the measurement result can be introduced through the interface			
5.6.1	Acting upon significant faults			
	Instrument made automatically inoperative; or			
	visible or audible indication until user takes action or fault disappears			
	Automatic instrument made inoperative automatically			
5.6.2	Indication check			
	Display check needed			
	Display check not needed			
	All elements of the indication are active and non-active long enough to be checked by the operator			
6.1	Software identification			
	software of a measuring instrument/component is clearly identified			
	– displayed on command, or			
	– displayed during operation, or			
	– displayed on start-up			
6.2	Correctness of algorithms and function			
	Measuring algorithms and functions of a measuring instrument appropriate and functionally correct			
	Possible to assess algorithms and functions either by metrological tests, software tests or software examination			
	No hidden or undocumented functions or parameters shall exist.			
6.3	Securing and protecting software, parameters, the measurand value and measurement data			
	Constructed in such a way that possibilities for unintentional, unauthorised, or intentional misuse are minimal			
6.3.1	Protected in such a way that evidence of any intervention (e.g. software updates) shall be available.			
	Protected against unauthorised modification, loading, or changes by swapping the memory device.			
6.3.2	Only clearly documented functions maybe activated by the user interface			
6.3.3	Legally relevant characteristics are protected such that evidence of intervention is available			
	Displaying or printing of the current legally relevant parameter settings is possible.			
6.3.4	Measurement data and the measurand value protected against modification			
	Fitted with a checking facility to ensure that if modification or corruption is detected, the measurement data and the measurand value are discarded or marked unusable.			
6.4	Not possible to fraudulently simulate legally relevant software for presenting measurement results, using easily available and manageable tools			

Checklist (continued)

Requirement (R129-1)		Passed	Failed	Remarks
6.5	Audit trail			
	Audit trail is protected			
	Not possible to delete or exchange audit trail			
	Content of audit trail shown or displayed on command			
	Storage device for audit trail has sufficient capacity for at least three successive verifications			
	Audit trails contains:			
	– success/failure of the update procedure			
	– time stamp of the event			
	– for software download: software identification of the installed version			
	– for software download: software identification of the previous installed version			
	– for parameter change: identification of the changed parameter			
	– for parameter change: the old and new values of the changed parameter			
6.6	Requirements for specific configurations of electronic instruments			
	Meets specific requirements for following technologies:			
6.7	Detection of significant defects			
	Software is made inoperative automatically or a visual or audible indication is provided automatically until user takes action when significant defect detected			
	All necessary documentation provided			
6.8	Support of durability protection			
	When durability is jeopardised the instrument starts measures to ensure further durability or become inoperable until the user takes action, or the problem disappears.			
	All necessary documentation provided			
Annex B.1	Specification and separation of legally relevant parts and specification of interfaces			
Annex B.1.1	Interfaces			
Annex B.1.2	Separation of components			
Annex B.1.3	Specification and separation of software parts			
Annex B.2	Shared indications;			
Annex B.3	Storage of data			
Annex B.4	Transmission of measured quantity values and/or measurement data			
Annex B.5	Operating systems			
Annex B.6	Software updates			
R 129-2, 1.1	Documentation			
	Submission accompanied by sufficient documentation, to ensure complete understanding of the construction and method of operation of the instrument including:			
	– drawings			
	– specifications			
	– photographs			
	– descriptions			
	Details of the measurement data contained in the memory and calculation methods provided			
	For electronic instruments, documentation includes:			
	– list of all electronic sub-assemblies with their essential characteristics			
	– description of electronic devices with drawings, diagrams and general software information explaining their construction and operation			

RESULT

PASS

FAIL

Use this page to detail remarks from the checklist: